



Ref. Certif. No.

US/8186/UL

IEC SYSTEM FOR CONFORMITY TESTING AND CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

# CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product  
Produit

Name and address of the applicant  
Nom et adresse du demandeur

Name and address of the manufacturer  
Nom et adresse du fabricant

Name and address of the factory  
Nom et adresse de l'usine

Rating and principal characteristics  
Valeurs nominales et caractéristiques principales

Trademark (if any)  
Marque de fabrique (si elle existe)

Model / Type Ref.  
Ref. de type

Additional information (if necessary)  
Information complémentaire (si nécessaire)

A sample of the product was tested and found to be in conformity with  
Un échantillon de ce produit a été essayé et a été considéré conforme à la

as shown in the Test Report Ref. No. which forms part of this Certificate  
comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

Switch Mode Power Supply

Elpac Electronics Inc.  
Elpac Power Systems, Div of  
1562 Reynolds Ave  
Irvine, CA 92614, USA

Elpac Electronics Inc.  
Elpac Power Systems, Div of  
1562 Reynolds Ave  
Irvine, CA 92614, USA

Elpac Electronics (Shenzhen) Inc.  
Minlida Ind Gardner, Sast Bldg, 5th Fl S  
Honghualing Industrial Zone  
Shenzhen, Nanshan, Xili Town  
Guangdong 518055 China

Input: 100-240 V ac, 50/60 Hz, 1.75 A  
Output min. 12 V dc, max. 24 V dc, max. 5.4 A  
Max. output power not to exceed 65 W



MW65XX and MSM65XX; where XX may be any numeric value which represents the output voltage.


The CB Test Report comprises 5 enclosures.

PUBLICATION EDITION

IEC 60601-1 (1988) Second Edition, with Amendment No. 1 (1991) and No. 2 (1995) with the exception of: Clause 36, Electromagnetic Compatibility, Clause 48, Biocompatibility, and Clause 52.1 (IEC 601-1-4), Programmable Electronic Systems. Inclusive of CENELEC Common Modifications. See Test Report for National Differences.

E149685-A11-CB-1

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



Underwriters Laboratories Inc. / Certification Programs Office  
333 Pfingsten Road, Northbrook, IL 60062-2096  
United States of America  
TEL INT\* 1-847-272-8800, Ext. 43008 FAX INT\* 1-847-272-9562  
email: jolanta.m.wroblewska@us.ul.com

Date:  
Issued: 2004 April 5

Signature:

Jolanta M. Wroblewska