



Ref. Certif. No.

US-TUVR-3698

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

Product
Produit

Medical switch mode power supply

Name and address of the applicant
Nom et adresse du demandeurElpac Electronics, Inc.
1562 Reynolds Avenue
Irvine, CA 92614 USAName and address of the manufacturer
Nom et adresse du fabricant

Same as applicant

Name and address of the factory
Nom et adresse de l'usineElpac Electronics (Shenzhen), Inc.
3F, SAST Building, MinLiDa Ind.
Garden, HongHuaLing Industrial Zone, XiLi Town, Nanshan District
Shenzhen City, Guangdong, PR China 318055Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} pageRatings and principal characteristics
Valeurs nominales et caractéristiques principalesInput: AC 100-240V~, 50/60Hz, 5A
Output: DC 12-48V, 20A Max (250W units), 28A Max (350W unit)
250W (w/convection cooling), 350W (with 26CFM min. Airflow)
Protection Class ITrademark (if any)
Marque de fabrique (si elle existe)Model / Type Ref.
Ref. De typeMSMP250XXF
(xx = 12 to 48)Additional information (if necessary, may also be reported
on page 2)Les informations complémentaires (si nécessaire, peuvent
être indiqués sur la 2^{ème} page)Complies with requirements as well as group and national differences
where applicable for AU, CA, DK, IL, KR and USA sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

PUBLICATION

EDITION

IEC 60601-1:1988+A1+A2

2nd Edition

As shown in the Test Report Ref. No. which forms part of
this CertificateComme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

30682009.001

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de CertificationTUV Rheinland
of North America

Date: May 31, 2007

Page 1 of 1

Dipl.-Ing. Matthias Heinze

Date:

Signature: