



Ref. Certif. No.

US/8018A/UL

IEC SYSTEM FOR CONFORMITY TESTING AND CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product
Produit

Switch Mode Power Supply

Name and address of the applicant
Nom et adresse du demandeur

Elpac Power Systems, Division of Elpac Electronics, Inc.
1562 Reynolds Ave.
Irvine, CA 92614, USA

Name and address of the manufacturer
Nom et adresse du fabricant

Elpac Power Systems, Division of Elpac Electronics, Inc.
1562 Reynolds Ave.
Irvine, CA 92614, USA

Name and address of the factory
Nom et adresse de l'usine

Elpac Electronics (Shenzhen) Inc.
Minlida Ind Gardner, SAST Bldg, 5th Fl S
Honghualing Industrial Zone
Shenzhen, Nanshan, Xili Town
Guangdong 518055, China

Rating and principal characteristics
Valeurs nominales et caractéristiques principales

Input: 100-240 V~, 50-60 Hz, 1.5 A
Output: min. 12 Vdc; max. 24 Vdc; 60 W

Trademark (if any)
Marque de fabrique (si elle existe)



Model / Type Ref.
Ref. de type

MSM60XX, where XX can be any number, 3600.

Additional information (if necessary)
Information complémentaire (si nécessaire)

The CB Test Report comprises 4 enclosures. The CB Test Certificate was amended on September 12, 2006 to add critical components.

PUBLICATION EDITION


A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1 (1988) Second Edition, with Amendment No. 1 (1991) and No. 2 (1995) with the exception of: Clause 36, Electromagnetic Compatibility, Clause 48, Biocompatibility, Clause 52.1 (IEC 601-1-4), Programmable Electronic Systems. See CB Test Report for National Differences.

as shown in the Test Report Ref. No. which forms part of this Certificate
comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

E149685-A10-CB-1

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



Underwriters Laboratories Inc. / Certification Programs Office
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* 1-847-272-8800, Ext. 43008 FAX INT* 1-847-313-3008
email: jolanta.m.wroblewska@us.ul.com

Date: Issued: 2004 February 24
Amended: 2006 September 12 (Am. 1)

Signature:

Jolanta M. Wroblewska