

Certificate



Certificate no.

CU 72080082 01

License Holder:

Elpac Electronics Inc.
Elpac Power Systems, Div. of
1562 Reynolds Avenue
Irvine CA 92614
USA

Manufacturing Plant:

Elpac Electronics Shenzhen
HongHuaLing Industrial Zone,
XiLi Town
5 F South,
SAST Building Min Lida Ind. Garden
518055 Nanshan District, Shenzhen City
China

Test report no.: USA-CC 30792209 002

Client Reference: Ricardo Felix

Tested to: UL 60601-1:2003 R6.03
CAN/CSA-C22.2 No. 601.1-M90

Certified Product: Medical Switch Mode Power Supply

License Fee - Units

Model Designation: 1) MSM80xxF (xx = 9-24)
2) MTA0yy0xxz (yy = 10-65,
xx = 09-24, z = A-Z)

Rated Voltage: AC 100-240V, 50/60Hz
Rated Current: 1) 1.5A (60W with convection cooling,
81.6W with 5 CFM Airflow)
2) 1.5A (50W max. with convect. cooling,
65W max. with 5 CFM Airflow)

Protection Class: I

contd.

Appendix: 1, 1-4

Licensed Test mark:



Signature

Dipl.-Ing. M. Raab
QA Certification Officer

Date of Issue
(day/mo/yr)

20/03/2008

Certificate



Certificate no.

CU 72080082 02

License Holder:

Elpac Electronics Inc.
Elpac Power Systems, Div. of
1562 Reynolds Avenue
Irvine CA 92614
USA

Manufacturing Plant:

Elpac Electronics Shenzhen
HongHuaLing Industrial Zone,
XiLi Town
5 F South,
SAST Building Min Lida Ind. Garden
518055 Nanshan District, Shenzhen City
China

Test report no.: USA-CC 30792209 002

Client Reference: Ricardo Felix

Tested to: UL 60601-1:2003 R6.03
CAN/CSA-C22.2 No. 601.1-M90

Certified Product: Medical Switch Mode Power Supply

License Fee - Units

contd.

Output Ratings DC: 1) 9-24V, 6.6A max. w/convection cooling
9-24V, 8.8A max. with 5 CFM Airflow
2) 9-24V, 5.6A max. w/convection cooling
9-24V, 7.2A max. with 5 CFM Airflow

Licensed Test mark:



Signature

A handwritten signature in black ink, appearing to read "M. Raap".

Dipl.-Ing. M. Raap
QA Certification Officer

Date of Issue

(day/mo/yr)

20/03/2008