

## **COVER PAGE FOR TEST REPORT**

Product Category:	Power Supplies, Medical and Dental
Product Category CCN:	QQHM2, QQHM8
Test Procedure:	Component Recognition
Product:	Switch Mode Power Supply
Model/Type Reference:	MW65XX and MSM65XX; where XX may be any numeric value which represents the output voltage and Model 3855.
Rating(s):	Input: All Models: 100-240 V ac 50/60 Hz 1.75 A  MW and MS Series: Output: Min. 12 V dc, Max. 24 V dc Max. 5.4 A Max. output power not to exceed 65 W  Model 3855 Output: 15 Vdc; 4.3 A, Max. 65 W
Standards:	UL 60601-1, First Edition (2003) CAN/CSA-C22.2 No.601.1-M90 with updates 1 and 2
Applicant Name and Address:	ELPAC ELECTRONICS INC 1562 REYNOLDS AVE IRVINE CA 92614
This Report includes the following parts, in addition to this cover page:	
<ol style="list-style-type: none"><li>1. Specific Technical Criteria</li><li>2. Clause Verdicts</li><li>3. Critical Components</li><li>4. Test Results</li><li>5. Enclosures</li></ol>	

Issue Date: 2004-04-02  
Amendment 1 2006-03-09

Page 2 of 2

Report Reference #


E149685-A11-UL-1

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Test Report By:



Bob Hilbrecht  
Staff Engineer  
Underwriters Laboratories Inc.

Reviewed By:



Amanda G. Pack  
Senior Project Engineer  
Underwriters Laboratories Inc.