



Ref. Certif. No.

US-TUVR-5056-A1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D' ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product
Produit

Medical Switch Mode Power Supply

Name and address of the applicant
Nom et adresse du demandeur

Elpac Electronics Inc.
Elpac Power Systems, Div. of
1562 Reynolds Avenue
Irvine, CA 92614 US

Name and address of the manufacturer
Nom et adresse du fabricant

same as applicant

Name and address of the factory
Nom et adresse de l'usine

Additional Information, see page 2

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Additional Information, see page 2

Trademark (if any)
Marque de fabrique (si elle existe)

Trademark

Model / Type Ref.
Ref. De Type

MTB0xx0yy (xx = 45-80; yy = 09-24)

Additional information (if necessary, may also be reported on page 2)
Les informations complémentaires (si nécessaire, peuvent être indiqués sur la 2^{ème} page)

Re-issue of Certificate US-TUVR-5056 due to 1st modification (change of Model / Type Ref.). See associated test report for details.

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la


IEC 60601-1 +Adm1+Amd2

2nd Edition (1988)

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

30991849.003

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



Date:

March 05, 2010

Page 1 of 2

Signature:

Martin Glagla



Ref. Certif. No.

US-TUVR-5056-A1

Additional Information:

Name and address of the factories:

- 1) Elpac Electronics (Shenzhen) Inc.
3F, SAST Building, MinLiDa Ind. Garden, HongHuaLing Industrial Zone,
XiLi Town, Nanshan District, Shenzhen City,
Guangdong 518055, P.R. China
- 2) Fabricators International Ltd.
3rd Row, Block A, Shang Yuan Road, 2nd Industrial Zone,
Quingxi Town, DongGuan City, Guangdong, P.R. China

Ratings and principal characteristics:

Input: AC 100-240V, 50/60Hz; 2A

Output Ratings: 1) DC 9-11.99V

7.2-5.42A max.; 65W max. (with convection cooling)

10.0-8.3A max.; 90W max. (with 5CFM min. airflow)


2) DC 12-24V

6.7-3.3A max.; 80W max. (with convection cooling)

8.3-4.18A max.; 100W max. (with 5CFM min. airflow)

Class I

This CB Test Certificate is issued by the National Certification body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



March 05, 2010

Page 2 of 2

Martin Glagla

Date:

Signature: