

COVER PAGE FOR TEST REPORT


Product Category:	Power Supplies, Medical and Dental
Product Category CCN:	QQHM2, QQHM8
Test Procedure:	Component Recognition
Product:	Switch Mode Power Supply
Model/Type Reference:	MSM75XX ;where XX can be any number between 12 and 28. XX represents the output voltage.
Rating(s):	Input: 100-240 V~, 50/60 Hz, 2.5 A Optional External cooling (26 CFM) Output: Min. 12 Vdc,; Max. 28 Vdc; 110 W Convection Cooling Output: Min. 12 Vdc; Max. 28 Vdc; 75 W
Standards:	UL 60601-1, First Edition (2003) CAN/CSA-C22.2 No.601.1-M90 with updates 1 and 2
Applicant Name and Address:	ELPAC ELECTRONICS INC 1562 REYNOLDS AVE IRVINE CA 92614
This Report includes the following parts, in addition to this cover page:	
<ol style="list-style-type: none">1. Specific Inspection Criteria2. Specific Technical Criteria3. Clause Verdicts4. Critical Components5. Test Results6. National Differences7. Enclosures	

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Test Report By:



Bob Hilbrecht
Staff Engineer
Underwriters Laboratories Inc.

Reviewed By:



Dean Klubnik
Senior Project Engineer
Underwriters Laboratories Inc.