

COVER PAGE FOR TEST REPORT

Product Category:	Power Supplies, Medical and Dental
Product Category CCN:	QQHM2, QQHM8
Test Procedure:	Component Recognition
Product:	Switch Mode Power Supply
Model/Type Reference:	MSM20XX; where XX can be any number.
Rating(s):	Model MSM20XX;
	Input: 100-240 V~, 50-60 Hz, 25W
	Output: Min. 12 V, Max. 24 Vdc; 20 W
Standards:	UL 60601-1, First Edition (2003)
Applicant Name and Address:	ELPAC ELECTRONICS INC ELPAC POWER SYSTEMS, DIV OF 1562 REYNOLDS AVE IRVINE CA 92614
This Report includes the following parts, in addition to this cover page:	
	1. Specific Inspection Criteria
	2. Specific Technical Criteria
	3. Clause Verdicts
	4. Critical Components
	5. Test Results
	6. National Differences
	7. Enclosures

This is to certify that representative samples of the products covered by this Test Report have been investigated by Underwriters Laboratories Inc. ('UL') in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of UL in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

UL authorizes the applicant to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Test Report By:

Reviewed By:

Mitchell McGarry
Project Engineer

Deborah Madsen
Staff Engineer
Underwriters Laboratories Inc.

Copyright © 2002 Underwriters Laboratories Inc.